

महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय, गोंदिया

Government Medical College, Gondia

कुंवर तिलकसिंह सामान्य रुग्णालय परिसर, नेहरु चौक, गोदिया, महाराष्ट्र-४४१६०१

Telephone No 07182-238090

Email-ID:- gmcgstore2015@gmail.com

Website: www.gmcgondia.in Ref. No./GMCG/PS/Qt/ ムナナ6

776 /2020

Date: 21 107 12020

Open Quotation Notice

For the supply of Instrument, Materials & Equipments

Dean Govt. Medical College, Gondia invites open quotation for the supply of INSTRUMENTS & EQUIPMENTS for various Departments from Manufacturer/Dealers. Detail list is enclosed. Interested suppliers/Manufacturer/dealer can send their quotation.

Quotation should be sent in sealed envelope (duly sealed by selling wax) by post or hand delivery so as to reach this office on or before 25 / 7-12020 upto 4 00 pm Date & timing of the opening of quotation 15 / 7-12020 at pm at college council Hall. Representative of the supplier may attend the meeting by submitting authorization letter.

- 1. The word "Quotation" for Instruments, Materials & Equipments and last Date 281 742020 Should be clearly written on cover.
- 2. Quotation, which received late, will not be accepted under any circumstances.
- 3. Rate should be for free delivery at Govt. Medical College, Gondia premises only. Rate should be quoted including all charges. Rates must be quote on letter head along with Pan and GST Number.
- 4. Quoted goods should be strictly accordingly to the specification mentioned in the list. Make Model of items should be specifically stated in quotation and catalogue /leaflets etc. should be submit along with the quotation.
- 5. Catalogue/Leaflet is must quotation received and the price of the same will not be considered on leaflets/Catalogue.
- 6. The concern Head of the Department can ask for demo of the lowest price model and it will be binding for the supplier to give the demo in stipulated period (i.e. within 4 days) at Gondia or nearby place. If the supplier fails to give the demonstration of the same model in stipulated period then his quotation will be treated as cancelled.
- 7. After placing the order to lowest one, the order will have to execute in full within the stipulated time if the supplier fail to comply, within stipulated period after giving the order may be cancelled and he will be declared defaulter and his any offer will not be consider in future.
- 8. The undersigned reserved the rights to accept or reject any or all quotation without giving any reason.
- 9. The quotation should be sent in the name of the Dean, Govt. Medical College, Gondia.
- 10. For the CMP purpose (unregistered) supplier should submit copy of Pan Card, Cancelled Check, One photograph, & Bank Details.
- 11. All inclusive warranty period should be two years from the date of commissioning of equipments.
- 12. The supplier should deliver the Instruments & Equipments in the concerned department of the institution of the institution give demonstration free of cost and staff of the department if required.

Government Medical College, Gondia

Government Medical College, Gondia List for quotation Name of Item with detail Specification

	Name of Item				
01	Colony counter				
	Specification				
	1) it should constructed with heavy guage mild steel & finished with powder				
	Contact analysis = 1st				
	2) It should have counting plate with illuminated standard wolfhugel ruling is fitted with magnifying glass on front slope. 3) Each touch of marking pencil in the colony count by 1 & the flowing LED digital depicts the colony count.				
	4) It should have digit diplay & maximum colony count possible upto 9999.				
	5) It should have 3 fluroscent light.6) It should be provided with manual reset of digital display.				
	7) It should have ON/OF switch, counting plate, 100 mm Dia lens, cord & plug.				
	8) It should work on 110V to 220 Volts A.C.				
02	Coagulomiter				
	Specification:- Fully automated, open random access, 12 detector channel Blood Coagulation				
	Analyzer having clotting, chromogenic and immunological assay channels. Clotting detection methods by recording change of viscosity and/or change of light transmitability and permitting both chromogenic and immunogenic assays Parameters				
	Provision fofr simultaneous random analysis for t least 15 parameters such as PT,				
	APTT, Fibrinogen, Protein S, Protein C, APC resistance, ProC, Antithrombin III,				
	Heparin, Plasminogen Activator inhibitor, Fibrin Monomer, D-Dimer Plus, von Willebrnd Antigen, Lupus Anticoagulant, and coagulation factors VIII & IX				
	Should have throughput of at least 100 tests per hour for PT.				
	Should be provided with a compatible printer.				
	Should have a intra assay reproducibility of:				
	Normal Patient				
	PT<2.0% PT < 2.0%				
	APTT < 2.0% Fibrinogen < 4% Fibrinogen < 5%				
	Fibrinogen < 4% Fibrinogen < 5% Additional Accessories required : 1) All standard accessories 1 set extra.				
	2) Compatible Onling UPS with 1 hour back up.				
(consumables : Quote list price of All consumables for 1000				
	tests.				
	be quoted separately				
03	Requirements at installation : To be installed by the Company				
5	Nebulizer (autoclavable with compressed base) Specification :-				
Ň	Model: NE-C28				
	Type: Non-heating compressor actuality				
	Power Consumption : 230V, 50Hz model ; 230V, 60 Hz Model 138VA				
F	Destination (without cap)				
	Medication Cup Capacity: *MMD approximately 5 um (based on EN13544-1)				

Appropriate Medication

Quantities:

2-7 ml (cc)

Operating Temperature/ **Humidity**:

Storage Temperature/

Humidity:

-20 C to +60 C, 10% to 95% RH Approx. 1.9kg (compressor only)

+10 C to +40 C, 30% to 85% RH

Weight: Dimensions:

Contents:

Approx. 170mm (w) x 103mm (h) x 182mm (d) Compressor, Nebulizer Kit, Air Tube, 5 pcs

Replacement Filters, Mouthpiece, Adult Mask,

Carrying Bag, Instruction Manual.

*MMD= Mas Median Diameter

Notes:

Subject to technical modification without prior notice.

Please note that specifications may very with medication type used.

Do not use the device where it may be exposed to flammable gas.

This unit confirms to EMC Standard IEC60601-1-2:2001. However, if it is used together with other medical devices or electrical equipment, they may influence the operationms of one of devices. Please follow any instructions in

the manuals and use all devices correctly.

04 **Dental Chair Automatic**

Specification:-

Luxurious Upholstery Design with Sturdy and Broad Base

Synchronize movement of Backrest & Seatproviding Trendelenburg chair position for syncope chair Powered by Linak Motor from Denmark

" Press Button Articulated adjustable head rest for Adult and Child Position.

" Broad Backrest Providing Extreme Comfort To Patient Single Lever Protection From Air, Water and Power Input

"Sofa Pattern Leather Upholstery Optional

" Option For Both Right And Left Handed Dentistry

DELIVERY UNIT

" All Tubings concealed at the base Of The Chair.

" Feather Touch Chair Movement Controls by Remote, Main Panel, Assistant

Tray and Foot Wwitch All Air and Water Tubes are Made in USA

" 9 Open programs for Chair Positions with two Fixed Zero and Gargle positions " 180 degree Movable Spittoon for Geriatric Patient Double Air-Rotor Points with Micrl motor Built In Terminal

Compactable Airotor Terminal Holder To Prevent Hand pieces From Falling

Two Three way Syringe, Operator and assistant " Pneumatic Locking Hinge

" Both Hot and Cold Water Facility

" LED Based Sensor operating Light With 360

Degree Rotation Movement with Varying Intensity

Government Medical College, Gondia

शासकीय वैद्यकीय महाविद्यालय, गोंदिया Government Medical College, Gondia

Sr. No.	Name of Machinery	Qty.	Dept.
1	Colony counter	01	Microbiology
2	Coagulomiter	01	Pathology
3	Nebulizer (autoclavable with compressd base)	01	Medicine
4	Dental Chair Automatic	01	Dental

Dean Government Medical College, Gondia